

09-19-05

JFW



Please type a plus sign (+) inside this box →


+

PTO/SB/21 (05-03)

Approved for use through 04/30/2003. OMB 0651-0031

U.S. Patent and Trademark Office; U.S. DEPARTMENT OF COMMERCE

Under the Paperwork Reduction Act of 1995, no persons are required to respond to a collection of information unless it displays a valid OMB control number.

<h1 style="text-align: center;">TRANSMITTAL FORM</h1> <p style="text-align: center;">(to be used for all correspondence after initial filing)</p>		Application Number	10/618,565
		Filing Date	July 11, 2003
		First Named Inventor	DECLERCK, JÉRÔME MARIE JOSEPH
		Group Art Unit	2621
		Examiner Name	COUSO, YON JUNG
Total Number of Pages in This Submission		Attorney Docket Number	KEMP-009(SP)
ENCLOSURES (check all that apply)			
<input type="checkbox"/> Fee Transmittal Form <input type="checkbox"/> Fee Attached <input type="checkbox"/> Amendment / Reply <input type="checkbox"/> After Final <input type="checkbox"/> Affidavits/declaration(s) <input type="checkbox"/> Extension of Time Request <input type="checkbox"/> Express Abandonment Request <input type="checkbox"/> Information Disclosure Statement <input checked="" type="checkbox"/> Certified Copy of Priority Document: GB0216854.0 <input type="checkbox"/> Response to Missing Parts/Incomplete Application <input type="checkbox"/> Response to Missing Parts under 37 CFR 1.52 or 1.53	<input type="checkbox"/> Assignment Papers (for an Application) <input type="checkbox"/> Drawing(s) <input type="checkbox"/> Licensing-related Papers <input type="checkbox"/> Petition <input type="checkbox"/> Petition to Convert to a Provisional Application <input type="checkbox"/> Power of Attorney, Revocation Change of Correspondence Address <input type="checkbox"/> Terminal Disclaimer <input type="checkbox"/> Request for Refund <input type="checkbox"/> CD, Number of CD(s)	<input type="checkbox"/> After Allowance Communication to Group <input type="checkbox"/> Appeal Communication to Board of Appeals and Interferences <input type="checkbox"/> Appeal Communication to Group (Appeal Notice, Brief, Reply Brief) <input type="checkbox"/> Proprietary Information <input type="checkbox"/> Status Letter <input checked="" type="checkbox"/> Return postcard	Remarks
SIGNATURE OF APPLICANT, ATTORNEY, OR AGENT			
Signing Attorney/Agent (Reg. No.)	CAROL M. LASALLE, 39,740 BOZICEVIC, FIELD & FRANCIS, LLP		
Signature			
Date	September 16, 2005		

EXPRESS MAIL LABEL NO. EV 687 633 402 US

This collection of information is required by 37 CFR 1.5. The information is required to obtain or retain a benefit by the public which is to file (and by the USPTO to process) an application. Confidentiality is governed by 35 U.S.C. 122 and 37 CFR 1.14. This collection is estimated to 12 minutes to complete, including gathering, preparing, and submitting the completed application form to the USPTO. Time will vary depending upon the individual case. Any comments on the amount of time you require to complete this form and/or suggestions for reducing this burden, should be sent to the Chief Information Officer, U.S. Patent and Trademark Office, U.S. Department of Commerce, P.O. Box 1450, Alexandria, VA 22313-1450. DO NOT SEND FEES OR COMPLETED FORMS TO THIS ADDRESS. SEND TO: Commissioner for Patents, P.O. Box 1450, Alexandria, VA 22313-1450.

If you need assistance in completing the form, call 1-800-PTO-9199 and select option 2.



The Patent Office
Concept House
Cardiff Road
Newport
South Wales
NP10 8QQ

, the undersigned, being an officer duly authorised in accordance with Section 74(1) and (4) of the Deregulation & Contracting Out Act 1994, to sign and issue certificates on behalf of the Comptroller-General, hereby certify that annexed hereto is a true copy of the documents as originally filed in connection with the patent application identified therein.

In accordance with the Patents (Companies Re-registration) Rules 1982, if a company named in this certificate and any accompanying documents has re-registered under the Companies Act 1985 with the same name as that with which it was registered immediately before re-registration save for the substitution as, or inclusion as, the last part of the name of the words "public limited company" or their equivalents in Welsh, references to the name of the company in this certificate and any accompanying documents shall be treated as references to the name with which it is so re-registered.

In accordance with the rules, the words "public limited company" may be replaced by p.l.c., plc, L.C. or PLC.

Re-registration under the Companies Act does not constitute a new legal entity but merely subjects the company to certain additional company law rules.

Signed 

Dated 1 September 2003

**CERTIFIED COPY OF
PRIORITY DOCUMENT**



1/77

Request for grant of a patent

(See the notes on the back of this form. You can also get an explanatory leaflet from the Patent Office to help you fill in this form)

The Patent Office

Cardiff Road
Newport
South Wales
NP10 8QQ

1. Your reference	N.86153 MN		
2. Patent application number (The Patent Office will fill in this part)	0216854.0 19 JUL 2002		
3. Full name, address and postcode of the or of each applicant (underline all surnames)	MIRADA SOLUTIONS LIMITED Oxford Centre for Innovation Mill Street Oxford OX2 0JX United Kingdom		
Patents ADP number (if you know it)			
If the applicant is a corporate body, give the country/state of its incorporation	United Kingdom		
4. Title of the invention	REGISTRATION OF MULTI-MODALITY DATA IN IMAGING		
5. Name of your agent (if you have one)	J.A. KEMP & CO.		
"Address for service" in the United Kingdom to which all correspondence should be sent (including the postcode)	14 South Square Gray's Inn London WC1R 5JJ		
Patents ADP number (if you know it)			
6. If you are declaring priority from one or more earlier patent applications, give the country and the date of filing of the or of each of these earlier applications and (if you know it) the or each application number	Country	Priority application number (if you know it)	Date of filing (day / month / year)
7. If this application is divided or otherwise derived from an earlier UK application, give the number and the filing date of the earlier application	Number of earlier application	Date of filing (day / month / year)	
8. Is a statement of inventorship and of right to grant of a patent required in support of this request? (Answer 'Yes' if: a) any applicant named in part 3 is not an inventor, or b) there is an inventor who is not named as an applicant, or c) any named applicant is a corporate body. See note (d))	Yes		

Patents Form 1/77

9. Enter the number of sheets for any of the following items you are filing with this form. Do not count copies of the same document

Continuation sheets of this form

Description	9
Claim(s)	3
Abstract	1
Drawing(s)	5

10. If you are also filing any of the following, state how many against each item.

Priority documents

Translations of priority documents


Statement of inventorship and right to grant of a patent (Patents Form 7/77)

Request for preliminary examination and search (Patents Form 9/77)

1

Request for substantive examination (Patents Form 10/77)

Any other documents (please specify)

11. I/We request the grant of a patent on the basis of this application.
Signature  Date 19 July 2002
J.A. KEMP & CO.

12. Name and daytime telephone number of person to contact in the United Kingdom
NICHOLLS, Michael John
020 7405 3292

Warning

After an application for a patent has been filed, the Comptroller of the Patent Office will consider whether publication or communication of the invention should be prohibited or restricted under Section 22 of the Patents Act 1977. You will be informed if it is necessary to prohibit or restrict your invention in this way. Furthermore, if you live in the United Kingdom, Section 23 of the Patents Act 1977 stops you from applying for a patent abroad without first getting written permission from the Patent Office unless an application has been filed at least 6 weeks beforehand in the United Kingdom for a patent for the same invention and either no direction prohibiting publication or communication has been given, or any such direction has been revoked.

Notes

- If you need help to fill in this form or you have any questions, please contact the Patent Office on 08459 500505.
- Write your answers in capital letters using black ink or you may type them.
- If there is not enough space for all the relevant details on any part of this form, please continue on a separate sheet of paper and write "see continuation sheet" in the relevant part(s). Any continuation sheet should be attached to this form.
- If you have answered 'Yes' Patents Form 7/77 will need to be filed.
- Once you have filled in the form you must remember to sign and date it.
- For details of the fee and ways to pay please contact the Patent Office.

REGISTRATION OF MULTI-MODALITY DATA IN IMAGING

The present invention relates to the registration of images of different modalities, in particular so that such images may be displayed together, accurately
5 superposed upon one another.

There are many fields in which it is useful to image a subject using different modalities. For instance, one modality might provide detailed structural information about the subject, for instance an x-ray image or a magnetic resonance image, while another modalities might provide information about different structures not visible in
10 the first modality, or information about functions occurring within the subject, such as by the introduction into the subject of a radioactive marker. While such different modality images can be considered side-by-side by someone trying to use the information given by the two different modalities, it is often useful to display the images in superposition one upon the other. It is clearly necessary for the
15 superposition to be accurate, in other words for areas representing a particular position in the subject in one image to be accurately positioned in registration with corresponding areas in the other image. The process of achieving this alignment is known as "registration". It is particularly useful because the display of the superposed images allows the information from the different modalities to be
20 interpreted very easily by a user. For example, the information about function from one modality can be related accurately to the detailed structural information from another modality.

A variety of techniques for registration of images of different modalities, particularly in the medical imaging field, have been proposed. For example, detailed
25 anatomical information about the structure of the body can be obtained from traditional x-ray images. Information about the metabolic function in the body can be obtained from different modalities, such as nuclear medicine. In a typical technique a radioactive marker is fixed to a physiological tracer. The tracer is injected in the blood of the patient and fixes to the cells in the patient according to
30 the metabolic activity (consumption of oxygen, or glucose etc). A detector is used to

detect the disintegration of the radioactive marker and to provide a corresponding image whose intensities correspond to the amount of radioactive marker in each region. The results of several scans may be combined together in the process known as tomography to provide 3-D information. Typical nuclear medicine techniques
5 include photon coincidence detection (PET) and single photon emission computerised tomography (SPECT). Such images are typically called emission images. One particularly important application of them is to the detection of tumours in the body. Such tumours are prominent in an emission images because of the high metabolic activity in and around the tumour. In the accompanying drawings Figure 1,
10 shows an example of an SPECT image, with areas of higher metabolic activity shown in lighter colours.

For some years such emission images have been acquired simultaneously with another image, called a transmission image, such as a single photon transmission computerised tomography (SPTCT) image, which is obtained by
15 placing a source of radiation on the opposite side of the subject's body from the detector. This provides an image which provides information regarding the attenuation and scattering characteristics of the subject's body. Figure 2 shows an example of such a transmission image. This transmission image is used to correct the emission image for the different attenuation and scattering occurring in different
20 areas of the body. The need for such correction is illustrated by Figures 3 and 4 of the accompanying drawings. The problem is that in an emission image, photons emitted from different parts of the body, e.g. from different depths or from different structures, will undergo different attenuation and scattering. This will lead to varying intensities in the image which are not related to the metabolic function in the body.
25 The transmission image effectively provides information about the differing attenuation and scattering and allows correction of the emission image. Figure 3 illustrates an emission image of a uniform cylindrical phantom before correction. It can be seen that although the phantom is uniform, and thus should have a uniform intensity, in fact the central areas are darker because of greater attenuation. Figure 4
30 illustrates such an image corrected by means of a transmission image. The

uniformity is restored.

A problem with nuclear medicine images, such as emission images, is that while they give good information about function, they do not give very good information about the structure of the subject. In particular, the exact location of regions of high metabolic function cannot be accurately determined. This is because the emission image does not show much structural detail. Many other imaging modalities reveal detailed structure, but obviously not the functional information of the nuclear medicine images. However, the lack of common information, i.e. the fact that the nuclear medicine images do not include detailed structure, and the fact that the other modality images do not include functional information, means that using the information from the two images, e.g. by matching the two images in order to register them accurately, is difficult. Figure 5 of the accompanying drawings shows a typical detailed structural image obtained, in this case, by x-ray imaging.

US Patent 5,871,013 and US Patent 5,672,877 both disclose methods of registering functional nuclear medicine emission images with other modality images, such as x-ray images, by using the transmission image (e.g. Figure 2) as a stepping stone in the registration process. This takes advantage of the fact that the transmission image is customarily obtained on the same imaging equipment as the functional emission image and so is inherently registered (i.e. the same pixel address in the image frame corresponds to the same position in the body). Furthermore, the transmission image includes some structural information, such as the edge of the body and the edges of the lungs, which are also visible in the different modality structural image. Therefore the registration of the transmission image with the different modality structural image is easier than direct registration of the functional emission image with the other modality image. In both patents, by first deriving a mathematical relationship mapping between the transmission image and the other modality image, a mapping between the functional emission image and the other modality can also be obtained because the relationship between the transmission image and the functional emission image is already known. Thus registration of the functional emission image with the other modality image is achieved without any

matching process between the two images.

In accordance with the present invention there is provided:-

a method of registering images of different modalities, comprising:

taking a first image of a subject obtained by an imaging process of a first
5 modality;

taking a second image of the subject obtained by an imaging process of a
second modality, said second image having a known positional relationship with the
first image;

taking a third image of the subject obtained by an imaging process of a third
10 modality;

distinguishing between at least one area of interest and at least one other area
not of interest in the second image;

on the basis of said known positional relationship identifying said at least one
area of interest and other area not of interest in the first image;

15 registering the first and third images by an image matching process based on
said at least one area of interest identified in the first image.

The at least one other area not of interest may be image of background
outside the subject, and possibly, in medical images, areas such as the lung cavity
within the body. Preferably the image matching process is conducted by looking
20 only, or mainly, at the areas of interest. One way of achieving this is to set the
intensities of the areas not of interest to a constant value, e.g. zero, so that the second
image has been used, effectively, as a mask to mask out areas which are not of
interest. However the area not of interest may be used to an extent in the registration
process, and in this case the second modality image is being used to segment into the
25 two areas, rather than to exclude one of them.

The first and second images are preferably obtained on the same imaging
apparatus, thus providing a known positional relationship, e.g. by being inherently
registered. The first image may be an emission image in which intensity values are
related to function in the subject, such as a PET or SPECT image. The second image
30 may be a transmission image of the type mentioned above. The third image may be

an image providing detailed structural information, such as an x-ray image, magnetic resonance image or ultrasound image.

5 The step of registering the first and third images may comprise deriving a positional transformation which maps to each other areas identified in the matching process as corresponding to each other. The matching process may be based on intensity or edge detection or another of the known techniques for matching two images.

10 Particularly in the medical field, the second image may be used as explained above to correct the first image for attenuation, and the first image may be further processed as is conventional, e.g. by equalisation.

The invention may be embodied in a computer system for processing data sets encoding the images, and the invention extends to a computer program for executing the method on a programmed computer. The invention also extends to a computer program product carrying such a computer program.

15 The invention will be further described by way of example, with reference to the accompanying drawings in which:-

Figure 1 is a functional emission image of a human body;

Figure 2 is a transmission image corresponding to Figure 1;

Figure 3 is a functional emission image of a uniform cylindrical phantom;

20 Figure 4 is an emission image corresponding to Figure 3 corrected for scattering and attenuation;

Figure 5 is an x-ray image of a human body showing detailed structural information;

25 Figure 6 is a block diagram schematically showing the apparatus needed for the imaging process;

Figure 7 is a schematic diagram illustrating the apparatus for processing and displaying the image data;

Figure 8 is a flow diagram explaining one embodiment of the invention;

Figure 9 is a transmission image with the main edges outlined;

30 Figure 10 is a mask generated from Figure 9;

Figure 11 is an enhanced functional emission image of the human body;

Figure 12 shows the image of Figure 11 after masking based on a transmission image;

Figure 13 illustrates the display of a functional emission image registered
5 with a detailed structural image.

Figure 6 illustrates schematically the apparatus needed for imaging. A patient
1 lies on a moveable support 3 beneath a detector 5. The detector is suitable for
whichever imaging modality is chosen, for instance a gamma or scintillation detector
10 for SPECT imaging. The signals from the detector 5 are supplied to an imaging
control and processing unit 7 which produces image data for display on display 9.
For the production of a transmission image a radioactive source 11 is provided on the
opposite side of the patient from the detector 5. Image data may also be acquired
using multiple projections from different angles of placement of the detector 5 (and
15 source 11 if necessary).

Typically image data is processed by computer before displayed as
schematically illustrated in Figure 7. Data sets encoding the image are stored in a
data store 13 and processed by a detector 15 before being displayed on a computer
display 17.

20 Figure 8 illustrates the registration process in accordance with an embodiment
of the present invention.

Firstly, in step 100 the three different images are obtained, in this example
one being a functional emission image such as a SPECT image, one a transmission
image such as a SPTCT image and one a structural image such as an x-ray image. In
25 step 104 an enhancement process is carried out on the emission image. The emission
image has a noisy background, and some features are extremely bright in the image:
the kidneys, the bladder and the liver are very bright as they evacuate the surplus of
radioactive contrast agent. This makes a histogram of the intensities in the image
very irregular with a lot of low intensities, a lot of very high intensities and a few
30 intermediate intensities. In the enhancement process the intensities are adjusted so as

to attenuate the very bright intensities. Such enhancement may be, for example, a gamma correction, or a histogram equalization, or other process which enhances the separation of features in the image. These techniques are standard in computer vision, can be found in, for example, "Digital Image Processing", by Nick Efford,
5 Addison Wesley, ISBN 0-201-59623-7, which is herein incorporated by reference.

In step 106 the transmission image is segmented to distinguish between different areas of the body and background. The aim is to identify areas which are not of interest in the image. This is achieved by smoothing the transmission image and then detecting the most significant edges in the image. In this example the edge
10 detection may be first detecting points of maximum intensity gradient. The intensity at each of the detected points is then collected and plotted in a histogram. A threshold value, which is the intensity value most represented in the histogram (the mode) is then defined. This intensity value is therefore the modal intensity of those pixels which are on a detected edge. Then all of the pixels in the image are examined
15 and the image is separated into two regions using the above modal intensity value as a threshold. The largest connected component or components of these regions are then extracted. However, other edge detection methods may be used. A transmission image in which the edges have been detected and marked (as a light outline) is shown in Figure 9.

20 Sometimes data sets may be presented, or stored, which include the emission image before and after correction for scattering and attenuation (by the transmission image), but the transmission image has been discarded. In this situation a version of the transmission image can be obtained for use in the invention by using the emission image before and after correction, because the difference between the
25 before and after images and a knowledge of the process of correction allows derivation of the transmission image used in that process. This derived "transmission image" may then be subjected to the edge detection process 106 described above.

The segmentation process 106 segments the image into two areas: the body without the lungs, and dark areas (lungs and background). In one example the dark
30 areas (lungs and background) can then be regarded as a mask as indicated in step

108. A mask generated from Figure 9 is shown in Figure 10. This mask is then applied to the equalised emission image in step 110. The application of the mask to the image is relatively easy because of the inherent registration of the transmission and emission images. The result of applying such a mask to the enhanced emission
5 image of Figure 11 is shown in Figure 12. In essence this process corresponds to setting intensity values to zero (or another constant value) in the emission image of the lung and background areas identified in the transmission image. However, it should be noted that in fact the segmented transmission image is indicating which pixels are inside the body and which are outside. Sometimes it may be useful to use
10 pixels outside the body in the registration process, as it can make the method more robust. But the knowledge provided by the segmented transmission image as to which pixels are inside the region of interest and which are outside is still useful. The registration algorithm can take account of the region not of interest as a class of intensities describing a particular area, and in this case they could be identified in the
15 mask by setting their intensity value to 1, to include them, rather than 0 to exclude them.

The masked emission image is then available for registration with an image obtained by another modality, such as an x-ray image. The fact that large areas of the image (which are not of interest) have been masked out, or at least identified, makes
20 the relevant and useful information in the combined emission and transmission image more specific. Any of the known matching and registration processes may be used, for instance based on detection and comparison of intensities, or intensity distributions, or detection and comparison of edges or of geometric structures. For example methods based on matching intensities include the calculation of a statistical
25 measure based on a joint histogram of the target image and the masked emission image, and changing the transformation parameters to minimise a similarity criterion. Other matching techniques which may be used are described, for example in "Accurate Three-Dimensional Registration of CT, PET and/or MR Images of the Brain", by Pelizzari, C.A., et al., Journal of Computer Assisted Tomography, Volume
30 13, 1989; "MRI-PET Registration with Automated Algorithm" by Woods, R.P., et

al., Journal of Computer Assisted Tomography, Volume 17, 1993; "The Principle
Axis Transformation - A Method for Image Registration", by Albert, N.M., et al.,
Journal of Nuclear Medicine, Volume 31, 1990; "New Feature Points Based on
Geometrical Invariance for 3-D Image Registration", Research Report Number 2149
5 from the INRIA, Jean-Philippe Thirion; and "A survey of medical Image
Registration", by J.B. Antoine Maintz, M. Viergever, Medical Image Analysis, 2(1):
1-36, 1998, all of which are herein incorporated by reference.

Once corresponding areas in the masked emission image and in the other
modality (e.g. x-ray) image have been identified, a mapping transformation which
10 indicates which pixels in the frame of the emission image correspond to which pixels
in the frame of the other modality image is obtained. This allows the two images to
be displayed superposed on one another in step 114. Figure 13 illustrates such a
display of two registered images, one a masked emission image and one an x-ray
image. Comparison of this with Figure 5 shows that an area of high metabolic
15 activity (high content of radioactive marker) is positioned in the left-hand region
(marked by the two orthogonal lines) clearly visible as a lighter area in Figure 11.
Also, differing levels of radioactive marker are visible in the central region. Because
Figure 11 includes the anatomically detailed X-ray CT scan it is easy for a clinician
to locate the abnormality (e.g. tumour), for instance to plan surgery, radiotherapy or
20 other treatment.

CLAIMS

1. A method of registering images of different modalities, comprising:
taking a first image of a subject obtained by an imaging process of a first
5 modality;
taking a second image of the subject obtained by an imaging process of a
second modality, said second image having a known positional relationship with the
first image;
taking a third image of the subject obtained by an imaging process of a third
10 modality;
distinguishing between at least one area of interest and at least one other area
not of interest in the second image;
on the basis of said known positional relationship identifying said at least one
area of interest and other area not of interest in the first image;
15 registering the first and third images by an image matching process based on
said at least one area of interest identified in the first image.
2. A method according to claim 1, wherein said at least one other area not of interest
comprises image of background outside the subject.
- 20 3. A method according to claim 1 or 2, further comprising the step of setting the
image intensities of the identified at least one other area not of interest to a constant
value prior to conducting said matching process.
- 25 4. A method according to claim 3, wherein said constant value is zero or one.
5. A method according to any one of the preceding claims wherein the first and
second images are obtained on the same imaging apparatus thus providing said
known positional relationship.

6. A method according to claim 7, wherein the first and second images are inherently registered.

7. A method according to any one of the preceding claims wherein the first image is
5 an emission image in which intensity values are related to function in the subject.

8. A method according to any one of the preceding claims wherein the second image
is a transmission image obtained by transmitting imaging radiation through the
subject from one side to the other, the intensity values being related to attenuation
10 and scattering of the radiation by the structure of the subject.

9. A method according to any one of the preceding claims wherein the third image is
a detailed structural image of the subject.

15 10. A method according to any one of the preceding claims wherein the step of
registering the first and third images comprises deriving a positional transformation
mapping to each other areas identified in said matching process as corresponding to
each other.

20 11. A method according to any one of the preceding claims wherein the matching
process comprises matching intensities of said at least one area of interest of the first
image with areas in said third image to identify corresponding areas.

12. A method according to any one of the preceding claims wherein the intensities in
25 said first image are corrected for attenuation in the subject by means of the second
image before said matching process is conducted.

13. A method according to any one of the preceding claims wherein the intensities in
said first image subjected to an enhancement process before said matching process is
30 conducted.

14. A method according to any one of the preceding claims wherein at least one of the first and second images is a nuclear medicine image.

5 15. A method according to any one of the preceding claims wherein the first image is a nuclear medicine image showing the presence of a radioactive marker in the body of the subject.

16. A method according to any one of the preceding claims wherein the third image is a medical image.

10

17. A computer system comprising a data processor, a data storage means and a display, the data processor being adapted to process data in accordance with an executable program stored in the data storage means, wherein the executable program is adapted to execute the method of any one of the preceding claims on input data
15 representing said first, second and third images and to display the first and third images superposed in registration with each other on the display.

18. A computer program comprising program code means for executing on a computer the method of any one of claims 1 to 16.

20

19. A computer program product carrying the computer program of claim 18.

ABSTRACT

REGISTRATION OF MULTI-MODALITY DATA IN IMAGING

5 A method of registration of a functional emission image to another image,
e.g. a structural/anatomical image, is disclosed in which the emission image is first
processed to mask out regions of the background and lungs identified in a
corresponding transmission image. Only areas which are not masked out in the
emission image are matched to areas in the x-ray image. The x-ray image and
10 masked emission image may then be displayed in superposition.




Figure 1




Figure 2




Figure 3




Figure 4




Figure 5




Figure 9




Figure 10




Fig. 6





Fig. 7

Fig. 8



5/5




Figure 11




Figure 12




Figure 13